



ANNEXURE-F FOR Steam Generator

Project #	061-001-ED-159
Document #	061-001-159-RFQ-006
Prep/Chk. By	WA / MAR
Date:	14-05-2024
Rev No.	2

Project	Development of LPG Testing Laboratory at Taftan			
S. No.	Description	Quantity	Unit Price (USD)	Total (USD)
1	Lab scale steam generator, electrically operated, 06-08 Kg/h, fire or water tube, operating Temperature 150°C-170°C, operating pressure 100 Psig meeting ASME or ISO standards.	2		
Total (USD)				
Note:				
1	All prices shall be CNF Karachi			
2	All technical details (Operation & Maintenance) must be provided with the proposal.			
3	Quotation shall be provided with Two (2) years operational spares with guaranties.			
4	Transportation to site at Taftan Border (Pak.-Iran) border.			